

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112222\
 Data File : VL039933.D
 Acq On : 23 Nov 2022 1:59
 Operator : SY/AP
 Sample : N5767-02DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/23/2022
 Supervised By :Mahesh Dadoda 11/24/2022

Quant Time: Nov 23 04:54:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Tue Nov 22 03:45:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	936757	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.639	114	2678710	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.455	117	2376088	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	1829103	10.078	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	= 100.800%	
Target Compounds						
						Qvalue
9) Propene	2.674	41	18347	0.301	ppbv	100
15) Acetone	3.462	43	147097	1.184	ppbv	95
20) Methylene Chloride	3.912	84	34964m	0.581	ppbv	
26) Hexane	5.195	57	68021	0.683	ppbv #	35
53) Toluene	9.204	91	29499	0.114	ppbv	99
74) 1,2,4-Trimethylbenzene	16.101	105	41822m	0.149	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112222\
Data File : VL039933.D
Acq On : 23 Nov 2022 1:59
Operator : SY/AP
Sample : N5767-02DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/23/2022
Supervised By : Mahesh Dadoda 11/24/2022

Quant Time: Nov 23 04:54:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Nov 22 03:45:31 2022
Response via : Initial Calibration

