

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112420\
 Data File : VL036005.D
 Acq On : 24 Nov 2020 22:23
 Operator : SY/AP
 Sample : L4843-02DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVP-02DL

Manual Integrations
 APPROVED

MMDadoda
 11/27/2020 1:16:22 PM

Quant Time: Nov 26 14:28:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1385042	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	4857357	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	4493406	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	3300663	10.34	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds					Ovalue
9) Propene	2.99	41	607090m	10.000	ppbv
10) Heptane	8.09	43	61000	0.364	ppbv 95
13) Ethanol	3.60	45	5079m	0.386	ppbv
26) Hexane	5.66	57	309198	1.922	ppbv # 44
32) 1,1,1-Trichloroethane	6.48	97	281255	0.852	ppbv 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112420\
Data File : VL036005.D
Acq On : 24 Nov 2020 22:23
Operator : SY/AP
Sample : L4843-02DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVP-02DL

Manual Integrations
APPROVED

MMDadoda
11/27/2020 1:16:22 PM

Quant Time: Nov 26 14:28:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

