

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112420\
 Data File : VL036007.D
 Acq On : 24 Nov 2020 23:40
 Operator : SY/AP
 Sample : L4843-03DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVP-03DL

Manual Integrations
 APPROVED

MMDadoda
 11/27/2020 1:16:18 PM

Quant Time: Nov 26 14:38:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1357013	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	4755224	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	4444705	10.00	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	3353079	10.62	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.20%
Target Compounds						Ovalue
9) Propene	3.00	41	785226m	13.202	ppbv	
10) Heptane	8.10	43	157868	0.961	ppbv	100
26) Hexane	5.66	57	642599	4.077	ppbv	# 45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112420\
Data File : VL036007.D
Acq On : 24 Nov 2020 23:40
Operator : SY/AP
Sample : L4843-03DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVP-03DL

Manual Integrations
APPROVED

MMDadoda
11/27/2020 1:16:18 PM

Quant Time: Nov 26 14:38:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

