

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112420\
 Data File : VL036008.D
 Acq On : 25 Nov 2020 00:17
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 11/27/2020 1:16:14 PM

Quant Time: Nov 25 07:19:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1141625	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	3905005	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	3636439	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2773518	10.74	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	8033m	0.161	ppbv	
13) Ethanol	3.59	45	62910m	5.795	ppbv	
15) Acetone	3.83	43	108681	0.890	ppbv #	87
20) Methylene Chloride	4.32	84	867763	10.100	ppbv	98
26) Hexane	5.66	57	459485	3.465	ppbv #	45
52) Tetrachloroethene	11.19	164	89197	0.649	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112420\
Data File : VL036008.D
Acq On : 25 Nov 2020 00:17
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
11/27/2020 1:16:14 PM

Quant Time: Nov 25 07:19:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

