

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112723\
 Data File : VL040903.D
 Acq On : 27 Nov 2023 10:19
 Operator : SY/MD
 Sample : VL1127ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1127ABL01

Quant Time: Nov 28 04:36:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Nov 28 04:31:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.227	49	917733	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.681	114	2266336	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.484	117	1922215	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.790	95	1528484	10.156	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112723\
Data File : VL040903.D
Acq On : 27 Nov 2023 10:19
Operator : SY/MD
Sample : VL1127ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1127ABL01

Quant Time: Nov 28 04:36:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Nov 28 04:31:59 2023
Response via : Initial Calibration

