

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL112918\
 Data File : VL032839.D
 Acq On : 29 Nov 2018 18:00
 Operator : SY/AP
 Sample : J6088-02DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IAS-4DL

Manual Integrations
 APPROVED

apatel
 12/4/2018 8:38:45 AM

Quant Time: Nov 30 07:22:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1466789	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3665355	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3455517	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2625257	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chlorodifluoromethane	3.02	51	18867m	0.162	ppbv	
9) Propene	3.06	41	8610m	0.185	ppbv	
13) Ethanol	3.67	45	27662	1.815	ppbv	94
15) Acetone	3.92	43	207978	1.553	ppbv	100
19) Isopropyl Alcohol	4.06	45	22297m	0.250	ppbv	
20) Methylene Chloride	4.42	84	23257m	0.335	ppbv	
26) Hexane	5.79	57	33043	0.238	ppbv #	44
34) 2-Butanone	5.36	43	33093	0.174	ppbv	100
38) Trichloroethene	7.97	130	7032m	0.060	ppbv	
53) Toluene	9.96	91	68630m	0.225	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL112918\
Data File : VL032839.D
Acq On : 29 Nov 2018 18:00
Operator : SY/AP
Sample : J6088-02DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IAS-4DL

Manual Integrations
APPROVED

apatel
12/4/2018 8:38:45 AM

Quant Time: Nov 30 07:22:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

