

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL113018\
 Data File : VL032853.D
 Acq On : 30 Nov 2018 17:43
 Operator : SY/AP
 Sample : J6132-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DSV-01DL

Manual Integrations
 APPROVED

apatel
 12/4/2018 4:11:05 PM

Quant Time: Dec 03 14:52:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1502616	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3782804	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3624651	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2780746	10.26	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
4) Chloromethane	3.17	50	11828	0.179	ppbv #	82
9) Propene	3.04	41	1191823	25.000	ppbv	90
13) Ethanol	3.66	45	62910m	4.029	ppbv	
15) Acetone	3.91	43	269170m	1.962	ppbv	
26) Hexane	5.79	57	145564	1.022	ppbv #	41
27) Carbon Disulfide	4.62	76	892037	4.941	ppbv	97
30) Cyclohexane	7.26	84	40866	0.328	ppbv #	21
34) 2-Butanone	5.34	43	240915	1.228	ppbv	100
36) Benzene	7.01	78	193016	0.641	ppbv	97
53) Toluene	9.95	91	237549	0.756	ppbv	99
59) m/p-Xylene	13.13	91	108232m	0.279	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL113018\
Data File : VL032853.D
Acq On : 30 Nov 2018 17:43
Operator : SY/AP
Sample : J6132-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DSV-01DL

Manual Integrations
APPROVED

apatel
12/4/2018 4:11:05 PM

Quant Time: Dec 03 14:52:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

