

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL113018\
 Data File : VL032856.D
 Acq On : 30 Nov 2018 19:41
 Operator : SY/AP
 Sample : J6132-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DSV-02DL

Manual Integrations
 APPROVED

apatel
 12/4/2018 4:11:14 PM

Quant Time: Dec 03 15:00:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1519164	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3784008	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3561709	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2677067	10.05	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	358459	7.437	ppbv	93
13) Ethanol	3.66	45	86629	5.487	ppbv	94
15) Acetone	3.91	43	201320m	1.451	ppbv	
27) Carbon Disulfide	4.63	76	55219	0.303	ppbv #	72
30) Cyclohexane	7.25	84	23234m	0.185	ppbv	
34) 2-Butanone	5.34	43	181853	0.927	ppbv	99
36) Benzene	7.01	78	69443	0.230	ppbv	98
53) Toluene	9.96	91	65060m	0.207	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL113018\
Data File : VL032856.D
Acq On : 30 Nov 2018 19:41
Operator : SY/AP
Sample : J6132-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DSV-02DL

Manual Integrations
APPROVED

apatel
12/4/2018 4:11:14 PM

Quant Time: Dec 03 15:00:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

