

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL113018\
 Data File : VL032865.D
 Acq On : 1 Dec 2018 1:31
 Operator : SY/AP
 Sample : J6132-06DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DSV-05DL

Manual Integrations
 APPROVED

apatel
 12/4/2018 4:11:23 PM

Quant Time: Dec 03 16:34:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1362993	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3422943	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3273487	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2545085	10.39	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	877088	20.283	ppbv	94
13) Ethanol	3.67	45	36899	2.605	ppbv	93
15) Acetone	3.91	43	209623m	1.684	ppbv	
27) Carbon Disulfide	4.63	76	297347	1.816	ppbv #	94
34) 2-Butanone	5.34	43	428223	2.412	ppbv	99
36) Benzene	7.01	78	87019	0.319	ppbv	97
53) Toluene	9.96	91	54359	0.191	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL113018\
Data File : VL032865.D
Acq On : 1 Dec 2018 1:31
Operator : SY/AP
Sample : J6132-06DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DSV-05DL

Manual Integrations
APPROVED

apatel
12/4/2018 4:11:23 PM

Quant Time: Dec 03 16:34:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

