

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL113020\
 Data File : VL036040.D
 Acq On : 1 Dec 2020 00:56
 Operator : SY/AP
 Sample : L3510-01DL 2X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E053-1100DL

Quant Time: Dec 01 05:25:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL113020BRAIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Nov 30 16:37:42 2020
 QLast Update : Mon Nov 30 16:37:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.70	49	1188430	10.00	ppbv	0.00
3) 1,4-Difluorobenzene	7.22	114	4082349	10.00	ppbv	0.00
4) Chlorobenzene-d5	12.14	117	3807648	10.00	ppbv	0.00

System Monitoring Compounds

5) 1-Bromo-4-Fluorobenzene	14.46	95	3124168	10.69	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Bromomethane	3.50	94	1464161	18.566	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL113020\
Data File : VL036040.D
Acq On : 1 Dec 2020 00:56
Operator : SY/AP
Sample : L3510-01DL 2X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E053-1100DL

Quant Time: Dec 01 05:25:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL113020BRAIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Nov 30 16:37:42 2020
QLast Update : Mon Nov 30 16:37:42 2020
Response via : Initial Calibration

