

Method Path : Z:\V0ASRV\HPCHEM1\MSV0A L\METHODS\

Method File : VL113020BRAIR.M

Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSV0A_L Mon Nov 30 16:37:42 2020

Last Update : Mon Nov 30 16:37:42 2020

Response Via : Initial Calibration

Calibration Files

0.5 =VL036030.D 1 =VL036029.D 2 =VL036028.D

10 =VL036027.D 15 =VL036031.D =

	Compound	0.5	1	2	10	15	Avg	%RSD
1) I	Bromochloromethane	-----ISTD-----						
2) T	Bromomethane	0.666	0.691	0.687	0.647	0.627	0.664	4.10
3) I	1,4-Difluorobenzene	-----ISTD-----						
4) I	Chlorobenzene-d5	-----ISTD-----						
5) S	1-Bromo-4-Fluorob	0.757	0.765	0.762	0.766	0.790	0.768	1.65

(#) = Out of Range ### Number of calibration levels exceeded format ###