

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL113020\
Data File : VL036032.D
Acq On : 30 Nov 2020 16:54
Operator : SY/AP
Sample : VSTDICV010
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL113020

Quant Time: Nov 30 18:23:16 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL113020BRAIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Nov 30 16:37:42 2020
QLast Update : Mon Nov 30 16:37:42 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 T	Bromomethane	0.664	0.556	16.3	88	0.00
3 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
4 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
5 S	1-Bromo-4-Fluorobenzene	0.768	0.788	-2.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0