

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL113020\
Data File : VL036044.D
Acq On : 1 Dec 2020 10:04
Operator : SY/AP
Sample : VSTDCCC010
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Dec 01 14:19:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL113020BRAIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Nov 30 16:37:42 2020
QLast Update : Mon Nov 30 16:37:42 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	90	0.00
2 T	Bromomethane	10.000	9.789	2.1	90	0.00
3 I	1,4-Difluorobenzene	10.000	10.000	0.0	84	0.00
4 I	Chlorobenzene-d5	10.000	10.000	0.0	83	0.00
5 S	1-Bromo-4-Fluorobenzene	10.000	10.253	-2.5	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0