

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120319\
 Data File : VL034456.D
 Acq On : 3 Dec 2019 12:14
 Operator : SY/AP
 Sample : K6068-04DL2 100X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-4DL2

Manual Integrations
 APPROVED

apatel
 12/3/2019 5:04:04 PM

Quant Time: Dec 03 12:47:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov 19 14:15:35 2019
 QLast Update : Tue Nov 19 14:15:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1048003	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.21	114	2107638	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.15	117	2047075	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.48	95	1752486	10.18	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.80%
Target Compounds						Ovalue
38) Trichloroethene	7.88	130	588230	6.803	ppbv	99
52) Tetrachloroethene	11.28	164	233727m	2.564	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120319\
Data File : VL034456.D
Acq On : 3 Dec 2019 12:14
Operator : SY/AP
Sample : K6068-04DL2 100X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-4DL2

Manual Integrations
APPROVED

apatel
12/3/2019 5:04:04 PM

Quant Time: Dec 03 12:47:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov
QLast Update : Tue Nov 19 14:15:35 2019
Response via : Initial Calibration

