

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120319\
 Data File : VL034458.D
 Acq On : 3 Dec 2019 13:30
 Operator : SY/AP
 Sample : K6068-02DL2 1200X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-2DL2

Manual Integrations
 APPROVED

apatel
 12/3/2019 5:04:09 PM

Quant Time: Dec 03 14:02:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 19 14:02:10 2019
 QLast Update : Tue Nov 19 14:15:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	978905	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.21	114	1912964	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.15	117	1929720	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.49	95	1613732	9.94	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
31) cis-1,2-Dichloroethene	5.56	61	14443m	0.130	ppbv	
38) Trichloroethene	7.88	130	383865	4.892	ppbv	98
52) Tetrachloroethene	11.28	164	18231m	0.220	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120319\
Data File : VL034458.D
Acq On : 3 Dec 2019 13:30
Operator : SY/AP
Sample : K6068-02DL2 1200X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-2DL2

Manual Integrations
APPROVED

apatel
12/3/2019 5:04:09 PM

Quant Time: Dec 03 14:02:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov
QLast Update : Tue Nov 19 14:15:35 2019
Response via : Initial Calibration

