

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120320\
 Data File : VL036055.D
 Acq On : 3 Dec 2020 17:48
 Operator : SY/AP
 Sample : VL1203ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1203ABL01

Quant Time: Dec 04 05:21:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 04 03:52:25 2020
 QLast Update : Fri Dec 04 03:52:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1150292	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3984061	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3671762	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2756568	9.87	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	25196	0.202	ppbv	# 1
20) Methylene Chloride	4.33	84	22748	0.222	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120320\
Data File : VL036055.D
Acq On : 3 Dec 2020 17:48
Operator : SY/AP
Sample : VL1203ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1203ABL01

Quant Time: Dec 04 05:21:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 04 03:52:25 2020
QLast Update : Fri Dec 04 03:52:25 2020
Response via : Initial Calibration

