

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120320\
Data File : VL036058.D
Acq On : 3 Dec 2020 19:50
Operator : SY/AP
Sample : OC113020 CAN 10597
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC113020 CAN 10597

Quant Time: Dec 04 11:49:50 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120320AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Dec 04 03:52:25 2020
QLast Update : Fri Dec 04 03:52:25 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1013375	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3531825	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3320129	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2464549	9.76	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.60%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120320\
Data File : VL036058.D
Acq On : 3 Dec 2020 19:50
Operator : SY/AP
Sample : OC113020 CAN 10597
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC113020 CAN 10597

Quant Time: Dec 04 11:49:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 04 03:52:25 2020
QLast Update : Fri Dec 04 03:52:25 2020
Response via : Initial Calibration

