

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120418\
 Data File : VL032895.D
 Acq On : 5 Dec 2018 3:16
 Operator : SY/AP
 Sample : J6176-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MMSV-08DL

Quant Time: Dec 05 15:56:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 07:00:23 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1478967	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3604533	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3656276	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2654277	9.71	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	90773	1.935	ppbv	99
13) Ethanol	3.66	45	68067	4.429	ppbv	98
15) Acetone	3.90	43	467694	3.463	ppbv #	75
34) 2-Butanone	5.33	43	2047154	10.950	ppbv	99
51) 2-Hexanone	10.30	43	137444	0.731	ppbv #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120418\
Data File : VL032895.D
Acq On : 5 Dec 2018 3:16
Operator : SY/AP
Sample : J6176-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MMSV-08DL

Quant Time: Dec 05 15:56:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 07:00:23 2018
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

