

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
 Data File : VL032924.D
 Acq On : 7 Dec 2018 6:11
 Operator : SY/AP
 Sample : VSTDICC0.1
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

MMDadoda
 12/11/2018 10:08:20 AM

Quant Time: Dec 07 07:25:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 07 07:25:28 2018
 QLast Update : Fri Dec 07 05:41:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1283136	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3045142	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	2829263	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2129443	10.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.30	62	5960	0.103	ppbv	93
32) 1,1,1-Trichloroethane	6.62	97	21724m	0.092	ppbv	
35) Carbon Tetrachloride	7.14	117	24055m	0.106	ppbv	
38) Trichloroethene	7.96	130	7742m	0.081	ppbv	
52) Tetrachloroethene	11.37	164	7353m	0.082	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.85	83	19904m	0.093	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
Data File : VL032924.D
Acq On : 7 Dec 2018 6:11
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
12/11/2018 10:08:20 AM

Quant Time: Dec 07 07:25:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 05:41:55 2018
Response via : Initial Calibration

