

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
 Data File : VL032925.D
 Acq On : 7 Dec 2018 6:48
 Operator : SY/AP
 Sample : VSTDIC0.03
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

MMDadoda
 12/11/2018 10:08:24 AM

Quant Time: Dec 07 07:28:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 07 07:28:28 2018
 QLast Update : Fri Dec 07 05:41:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1237249	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2872198	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	2741725	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2034272	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.30	62	2176m	0.039	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	8337m	0.037	ppbv	
35) Carbon Tetrachloride	7.14	117	9357m	0.044	ppbv	
38) Trichloroethene	7.96	130	3326m	0.037	ppbv	
52) Tetrachloroethene	11.38	164	2641m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.84	83	8363m	0.040	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
Data File : VL032925.D
Acq On : 7 Dec 2018 6:48
Operator : SY/AP
Sample : VSTDIC0.03
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Manual Integrations
APPROVED

MMDadoda
12/11/2018 10:08:24 AM

Quant Time: Dec 07 07:28:28 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 05:41:55 2018
Response via : Initial Calibration

