

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
 Data File : VL032938.D
 Acq On : 7 Dec 2018 17:12
 Operator : SY/AP
 Sample : J6262-02DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2DL

Manual Integrations
 APPROVED

MMDadoda
 12/11/2018 10:09:02 AM

Quant Time: Dec 10 20:44:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 0
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1385706	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3471509	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	3266286	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2481296	10.28	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chlorodifluoromethane	3.02	51	164658	1.569	ppbv	97
9) Propene	3.04	41	27750m	0.669	ppbv	
13) Ethanol	3.65	45	948961	66.419	ppbv	97
15) Acetone	3.91	43	709517	5.965	ppbv	93
19) Isopropyl Alcohol	4.03	45	1598996	21.407	ppbv #	99
25) Ethyl Acetate	5.77	43	964925	3.434	ppbv #	90
29) Chloroform	5.85	83	25665m	0.111	ppbv	
34) 2-Butanone	5.34	43	24126	0.160	ppbv	95
36) Benzene	7.01	78	45864	0.180	ppbv #	92
38) Trichloroethene	7.96	130	13088m	0.123	ppbv	
52) Tetrachloroethene	11.38	164	48295m	0.520	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
Data File : VL032938.D
Acq On : 7 Dec 2018 17:12
Operator : SY/AP
Sample : J6262-02DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2DL

Manual Integrations
APPROVED

MMDadoda
12/11/2018 10:09:02 AM

Quant Time: Dec 10 20:44:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

