

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
 Data File : VL032944.D
 Acq On : 7 Dec 2018 21:11
 Operator : SY/AP
 Sample : J6262-07DL 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 12/11/2018 10:09:27 AM

Quant Time: Dec 08 03:58:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 0
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1284153	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3225569	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	3055066	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2336667	10.35	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.05	41	13672	0.356	ppbv	83
13) Ethanol	3.67	45	27283m	2.061	ppbv	
15) Acetone	3.93	43	101759	0.923	ppbv #	83
20) Methylene Chloride	4.41	84	37710	0.700	ppbv	97
26) Hexane	5.78	57	78423	0.652	ppbv #	51
53) Toluene	9.95	91	175095	0.696	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL120718\
Data File : VL032944.D
Acq On : 7 Dec 2018 21:11
Operator : SY/AP
Sample : J6262-07DL 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

MMDadoda
12/11/2018 10:09:27 AM

Quant Time: Dec 08 03:58:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

