

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120720\
 Data File : VL036069.D
 Acq On : 7 Dec 2020 15:36
 Operator : SY/AP
 Sample : VSTDIC0.03
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

MMDadoda
 12/10/2020 9:53:22 AM

Quant Time: Dec 07 16:27:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 07 13:39:30 2020
 QLast Update : Mon Dec 07 13:39:30 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.684	49	1331033	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.195	114	2941821	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.108	117	2604004	10.00	ppbv	# 0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.442	95	2114737	10.63	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.30%

Target Compounds					Qvalue
5) Vinyl Chloride	3.269	62	3849m	0.04	ppbv
32) 1,1,1-Trichloroethane	6.526	97	6409m	0.03	ppbv
35) Carbon Tetrachloride	7.031	117	5297m	0.03	ppbv
38) Trichloroethene	7.861	130	3300m	0.04	ppbv
52) Tetrachloroethene	11.237	164	3824m	0.05	ppbv
63) 1,1,2,2-Tetrachloroethane	13.703	83	6762m	0.03	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120720\
Data File : VL036069.D
Acq On : 7 Dec 2020 15:36
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
12/10/2020 9:53:22 AM

Quant Time: Dec 07 16:27:43 2020

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120720AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 07 13:39:30 2020

QLast Update : Mon Dec 07 13:39:30 2020

Response via : Initial Calibration

