

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120720\
 Data File : VL036095.D
 Acq On : 8 Dec 2020 10:57
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Dec 08 12:14:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 07 16:56:49 2020
 QLast Update : Mon Dec 07 16:56:49 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	1.165	1.055	9.4	94	0.00
3	Chlorodifluoromethane	1.141	1.210	-6.0	107	0.00
4	Chloromethane	0.560	0.576	-2.9	104	0.00
5 T	Vinyl Chloride	0.623	0.569	8.7	101	0.00
6 T	Bromomethane	0.322	0.312	3.1	101	0.00
7	Chloroethane	0.212	0.204	3.8	104	0.00
8 T	Dichlorotetrafluoroethane	1.295	1.310	-1.2	102	0.00
9 T	Propene	0.572	0.562	1.7	106	0.00
10 T	Heptane	1.460	1.446	1.0	102	0.00
11 T	Trichlorofluoromethane	1.370	1.368	0.1	100	0.00
12 T	1,1,2-Trichlorotrifluoroeth	0.931	0.869	6.7	94	0.00
13	Ethanol	0.079	0.086	-8.9	115	0.00
14 T	Bromoethene	0.382	0.376	1.6	99	0.00
15 T	Acetone	1.202	1.163	3.2	109	0.00
16 T	1,3-Butadiene	0.605	0.584	3.5	99	0.00
17	tert-Butyl alcohol	1.164	0.921	20.9	79	0.00
18 T	1,1-Dichloroethene	0.413	0.382	7.5	92	0.00
19 T	Isopropyl Alcohol	0.735	0.687	6.5	97	0.00
20 T	Methylene Chloride	0.492	0.351	28.7	92	0.00
21 T	Allyl Chloride	0.799	0.694	13.1	85	0.00
22 T	trans-1,2-Dichloroethene	0.414	0.388	6.3	93	0.00
23 T	Vinyl Acetate	1.600	1.556	2.8	97	0.00
24 T	1,1-Dichloroethane	1.125	1.075	4.4	99	0.00
25 T	Ethyl Acetate	2.401	2.381	0.8	100	0.00
26 T	Hexane	1.081	0.982	9.2	101	0.00
27 T	Carbon Disulfide	1.074	0.955	11.1	85	0.00
28 T	Methyl tert-Butyl Ether	1.536	1.380	10.2	91	0.00
29 T	Chloroform	1.419	1.472	-3.7	105	0.00
30 T	Cyclohexane	0.794	0.780	1.8	103	0.00
31 T	cis-1,2-Dichloroethene	0.982	1.020	-3.9	104	0.00
32 T	1,1,1-Trichloroethane	1.464	1.424	2.7	101	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
34 T	2-Butanone	0.663	0.697	-5.1	106	0.00
35 T	Carbon Tetrachloride	0.583	0.634	-8.7	97	0.00
36 T	Benzene	0.847	0.932	-10.0	105	0.00
37 T	1,2-Dichloroethane	0.491	0.572	-16.5	108	0.00
38 T	Trichloroethene	0.317	0.281	11.4	90	0.00
39 T	1,2-Dichloropropane	0.335	0.373	-11.3	103	0.00
40 T	1,4-Dioxane	0.105	0.124	-18.1	114	0.00
41 T	Tetrahydrofuran	0.394	0.416	-5.6	102	0.00
42 T	Bromodichloromethane	0.654	0.742	-13.5	102	0.00
43	Methyl Methacrylate	0.364	0.416	-14.3	104	0.00
44 T	2,2,4-Trimethylpentane	1.535	1.638	-6.7	103	0.00
45 T	t-1,3-Dichloropropene	0.374	0.434	-16.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120720\
 Data File : VL036095.D
 Acq On : 8 Dec 2020 10:57
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Dec 08 12:14:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 07 16:56:49 2020
 QLast Update : Mon Dec 07 16:56:49 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.468	0.532	-13.7	102	0.00
47 T	1,1,2-Trichloroethane	0.322	0.351	-9.0	104	0.00
48 T	Dibromochloromethane	0.486	0.538	-10.7	95	0.00
49 T	Bromoform	0.383	0.438	-14.4	94	0.00
50 T	4-Methyl-2-Pentanone	0.989	1.091	-10.3	104	0.00
51 T	2-Hexanone	0.773	0.887	-14.7	109	0.00
52 T	Tetrachloroethene	0.307	0.277	9.8	94	0.00
53 T	Toluene	0.952	1.035	-8.7	103	0.00
54 T	1,2-Dibromoethane	0.483	0.519	-7.5	102	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	0.415	0.424	-2.2	96	0.00
57 T	Chlorobenzene	0.702	0.723	-3.0	101	0.00
58 T	Ethyl Benzene	1.370	1.467	-7.1	106	0.00
59 T	m/p-Xylene	1.154	1.229	-6.5	105	0.00
60 T	o-Xylene	1.088	1.157	-6.3	107	0.00
61 T	Styrene	0.698	0.801	-14.8	108	0.00
62	Isopropylbenzene	1.612	1.671	-3.7	104	0.00
63 T	1,1,2,2-Tetrachloroethane	0.732	0.782	-6.8	111	0.00
64	n-propylbenzene	0.388	0.412	-6.2	103	0.00
65	tert-Butylbenzene	1.319	1.374	-4.2	105	0.00
66 T	Benzyl Chloride	0.329	0.423	-28.6	98	0.00
67	sec-Butylbenzene	1.847	2.008	-8.7	110	0.00
68 S	1-Bromo-4-Fluorobenzene	0.794	0.836	-5.3	97	0.00
69	p-Isopropyltoluene	1.469	1.592	-8.4	108	0.00
70	n-Butylbenzene	1.452	1.743	-20.0	119	0.00
71	2-Chlorotoluene	1.219	1.364	-11.9	112	0.00
72 T	4-Ethyltoluene	1.178	1.310	-11.2	110	0.00
73 T	1,3,5-Trimethylbenzene	1.125	1.232	-9.5	109	0.00
74 T	1,2,4-Trimethylbenzene	1.129	1.237	-9.6	110	0.00
75 T	1,3-Dichlorobenzene	0.565	0.612	-8.3	106	0.00
76 T	1,4-Dichlorobenzene	0.576	0.613	-6.4	106	0.00
77 T	1,2-Dichlorobenzene	0.540	0.598	-10.7	112	0.00
78 T	Hexachloro-1,3-Butadiene	0.433	0.359	17.1	85	0.00
79 T	Naphthalene	0.574	0.649	-13.1	92	0.00
80 T	Naphthalene,2-methyl-	0.128	0.258	-101.6#	89	0.00
81 T	1,2,4-Trichlorobenzene	0.384	0.370	3.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0