

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043308.D
 Acq On : 08 Dec 2025 11:11
 Operator : SY/MD
 Sample : QC120225 CAN 10297
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC120225 CAN 10297

Quant Time: Dec 09 00:09:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	151181	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	400309	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	315883	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	225215	9.727	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043308.D
Acq On : 08 Dec 2025 11:11
Operator : SY/MD
Sample : QC120225 CAN 10297
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC120225 CAN 10297

Quant Time: Dec 09 00:09:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

