

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043328.D
 Acq On : 09 Dec 2025 00:06
 Operator : SY/MD
 Sample : Q3797-03DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 01:26:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	136077	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	340593	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	259356	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	186262	9.798	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.000%
Target Compounds						
					Qvalue	
15) Acetone	1.832	43	14609	0.773	ppbv	97
38) Trichloroethene	4.535	130	11121	0.871	ppbv	91
52) Tetrachloroethene	8.209	164	2779m	0.229	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043328.D
Acq On : 09 Dec 2025 00:06
Operator : SY/MD
Sample : Q3797-03DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 01:26:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

