

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043330.D
 Acq On : 09 Dec 2025 01:11
 Operator : SY/MD
 Sample : Q3798-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Quant Time: Dec 09 01:51:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	135207	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	342039	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.868	117	262872	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	189370	9.828	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%
Target Compounds						
					Qvalue	
9) Propene	1.482	41	1528	0.192	ppbv #	1
15) Acetone	1.835	43	6355	0.338	ppbv	98
29) Chloroform	2.845	83	292025	11.567	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043330.D
Acq On : 09 Dec 2025 01:11
Operator : SY/MD
Sample : Q3798-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Quant Time: Dec 09 01:51:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

