

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120919\
 Data File : VL034469.D
 Acq On : 9 Dec 2019 18:06
 Operator : SY/AP
 Sample : K6195-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NS-1DL

Quant Time: Dec 10 04:13:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 19 14:15:35 2019
 QLast Update : Tue Nov 19 14:15:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	965207	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.20	114	1900049	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1908655	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1626854	10.13	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	118835	7.979	ppbv	90
15) Acetone	3.84	43	90638	0.656	ppbv	99
19) Isopropyl Alcohol	3.96	45	71165	0.803	ppbv #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120919\
Data File : VL034469.D
Acq On : 9 Dec 2019 18:06
Operator : SY/AP
Sample : K6195-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NS-1DL

Quant Time: Dec 10 04:13:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 19 14:15:35 2019
QLast Update : Tue Nov 19 14:15:35 2019
Response via : Initial Calibration

