

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL121620\
 Data File : VL036186.D
 Acq On : 17 Dec 2020 3:45
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 12/17/2020 3:56:59 PM

Quant Time: Dec 17 07:51:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 07 2020
 QLast Update : Mon Dec 07 16:56:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1468564	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2798031	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2554604	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2168343	10.69	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.90%

Target Compounds					Qvalue
15) Acetone	3.88	43	27325m	0.155	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL121620\
Data File : VL036186.D
Acq On : 17 Dec 2020 3:45
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
12/17/2020 3:56:59 PM

Quant Time: Dec 17 07:51:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL120720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 07 16:56:49 2020
Response via : Initial Calibration

