

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL122018\
 Data File : VL033025.D
 Acq On : 20 Dec 2018 14:47
 Operator : SY/AP
 Sample : VL1220ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1220ABL01

Quant Time: Dec 21 06:28:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Dec 07 09:35:17 2018
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1242367	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2861522	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2759923	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2051452	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
20) Methylene Chloride	4.41	84	21243	0.408	ppbv	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL122018\
Data File : VL033025.D
Acq On : 20 Dec 2018 14:47
Operator : SY/AP
Sample : VL1220ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1220ABL01

Quant Time: Dec 21 06:28:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 07 09:35:17 2018
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

