

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
 Data File : VL040960.D
 Acq On : 22 Dec 2023 15:38
 Operator : SY/MD
 Sample : QC120523 CAN 10428
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC120523 CAN 10428

Quant Time: Dec 23 01:21:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 22 23:57:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1043835	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	2531379	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.436	117	2163763	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.745	95	1589556	9.973	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.700%
Target Compounds						
					Qvalue	
15) Acetone	3.520	43	38093	0.304	ppbv #	75
20) Methylene Chloride	3.951	84	37882	0.849	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
Data File : VL040960.D
Acq On : 22 Dec 2023 15:38
Operator : SY/MD
Sample : QC120523 CAN 10428
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC120523 CAN 10428

Quant Time: Dec 23 01:21:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 22 23:57:16 2023
Response via : Initial Calibration

