

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
 Data File : VL040961.D
 Acq On : 22 Dec 2023 16:56
 Operator : SY/MD
 Sample : QC112823 CAN 10306
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC112823 CAN 10306

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/23/2023
 Supervised By :Mahesh Dadoda 12/26/2023

Quant Time: Dec 23 01:22:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Dec 22 23:57:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	1002087	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2394506	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.436	117	1996874	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.744	95	1499979	10.198	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.000%
Target Compounds						
						Qvalue
20) Methylene Chloride	3.951	84	13773m	0.322	ppbv	
52) Tetrachloroethene	10.590	164	40191m	0.637	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
Data File : VL040961.D
Acq On : 22 Dec 2023 16:56
Operator : SY/MD
Sample : QC112823 CAN 10306
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC112823 CAN 10306

Quant Time: Dec 23 01:22:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 22 23:57:16 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/23/2023
Supervised By : Mahesh Dadoda 12/26/2023

