

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
 Data File : VL040962.D
 Acq On : 22 Dec 2023 17:33
 Operator : SY/MD
 Sample : QC120523 CAN 10428
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC120523 CAN 10428

Quant Time: Dec 23 01:22:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 22 23:57:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1013105	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	2393980	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.439	117	1977242	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.745	95	1499031	10.293	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122223\
Data File : VL040962.D
Acq On : 22 Dec 2023 17:33
Operator : SY/MD
Sample : QC120523 CAN 10428
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC120523 CAN 10428

Quant Time: Dec 23 01:22:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 22 23:57:16 2023
Response via : Initial Calibration

