

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX010219\
 Data File : VX006871.D
 Acq On : 02 Jan 2019 10:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 03 00:19:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.457	0.402	12.0	86	0.00
3 P	Chloromethane	0.493	0.439	11.0	88	0.00
4 C	Vinyl Chloride	0.515	0.475	7.8#	91	0.00
5 T	Bromomethane	0.560	0.431	23.0#	83	0.00
6 T	Chloroethane	0.356	0.307	13.8	91	0.00
7 T	Trichlorofluoromethane	0.802	0.779	2.9	94	0.00
8 T	Diethyl Ether	0.319	0.309	3.1	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.471	0.474	-0.6	98	0.00
10 T	Methyl Iodide	0.674	0.708	-5.0	93	0.00
11 T	Tert butyl alcohol	0.120	0.107	10.8	86	0.00
12 CM	1,1-Dichloroethene	0.447	0.426	4.7#	94	0.00
13 T	Acrolein	0.081	0.090	-11.1	109	0.00
14 T	Allyl chloride	0.752	0.763	-1.5	98	0.00
15 T	Acrylonitrile	0.267	0.268	-0.4	95	0.00
16 T	Acetone	0.248	0.251	-1.2	100	0.00
17 T	Carbon Disulfide	1.182	1.006	14.9	85	0.00
18 T	Methyl Acetate	0.728	0.792	-8.8	105	0.00
19 T	Methyl tert-butyl Ether	1.483	1.584	-6.8	102	0.00
20 T	Methylene Chloride	0.512	0.499	2.5	99	0.00
21 T	trans-1,2-Dichloroethene	0.495	0.455	8.1	94	0.00
22 T	Diisopropyl ether	1.376	1.484	-7.8	103	0.00
23 T	Vinyl Acetate	1.125	1.248	-10.9	105	0.00
24 P	1,1-Dichloroethane	0.821	0.847	-3.2	104	0.00
25 T	2-Butanone	0.348	0.344	1.1	95	0.00
26 T	2,2-Dichloropropane	0.634	0.703	-10.9	107	0.00
27 T	cis-1,2-Dichloroethene	0.546	0.540	1.1	95	0.00
28 T	Bromochloromethane	0.378	0.364	3.7	97	0.00
29 T	Tetrahydrofuran	0.229	0.220	3.9	92	0.00
30 C	Chloroform	0.850	0.856	-0.7#	98	0.00
31 T	Cyclohexane	0.746	0.682	8.6	89	0.00
32 T	1,1,1-Trichloroethane	0.718	0.754	-5.0	97	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.510	5.4	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.316	0.333	-5.4	91	0.00
36 T	1,1-Dichloropropene	0.414	0.432	-4.3	92	0.00
37 T	Ethyl Acetate	0.431	0.463	-7.4	97	0.00
38 T	Carbon Tetrachloride	0.402	0.465	-15.7	100	0.00
39 T	Methylcyclohexane	0.545	0.510	6.4	84	0.00
40 TM	Benzene	1.259	1.318	-4.7	93	0.00
41 T	Methacrylonitrile	0.229	0.267	-16.6	98	0.00
42 TM	1,2-Dichloroethane	0.429	0.454	-5.8	97	0.00
43 T	Isopropyl Acetate	0.656	0.780	-18.9	102	0.00
44 TM	Trichloroethene	0.381	0.392	-2.9	94	0.00
45 C	1,2-Dichloropropane	0.306	0.341	-11.4#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX010219\
 Data File : VX006871.D
 Acq On : 02 Jan 2019 10:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 03 00:19:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.225	0.243	-8.0	95	0.00
47 T	Bromodichloromethane	0.374	0.449	-20.1#	103	0.00
48 T	Methyl methacrylate	0.334	0.390	-16.8	98	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	75	0.00
50 S	Toluene-d8	1.192	1.223	-2.6	89	0.00
51 T	4-Methyl-2-Pentanone	0.417	0.471	-12.9	95	0.00
52 CM	Toluene	0.813	0.864	-6.3#	93	0.00
53 T	t-1,3-Dichloropropene	0.401	0.518	-29.2#	105	0.00
54 T	cis-1,3-Dichloropropene	0.440	0.550	-25.0#	102	0.00
55 T	1,1,2-Trichloroethane	0.333	0.370	-11.1	98	0.00
56 T	Ethyl methacrylate	0.470	0.542	-15.3	95	0.00
57 T	1,3-Dichloropropane	0.536	0.587	-9.5	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.277	-10.8	97	0.00
59 T	2-Hexanone	0.319	0.363	-13.8	96	0.00
60 T	Dibromochloromethane	0.307	0.409	-33.2#	107	0.00
61 T	1,2-Dibromoethane	0.354	0.396	-11.9	96	0.00
62 S	4-Bromofluorobenzene	0.437	0.446	-2.1	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.396	0.422	-6.6	98	0.00
65 PM	Chlorobenzene	1.058	1.113	-5.2	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.330	0.412	-24.8#	104	0.00
67 C	Ethyl Benzene	1.729	1.855	-7.3#	95	0.00
68 T	m/p-Xylenes	0.682	0.736	-7.9	94	0.00
69 T	o-Xylene	0.674	0.722	-7.1	93	0.00
70 T	Styrene	1.066	1.172	-9.9	93	0.00
71 P	Bromoform	0.251	0.334	-33.1#	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.421	3.625	-6.0	94	0.00
74 T	N-amyl acetate	1.224	1.455	-18.9	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.017	1.110	-9.1	95	0.00
76 T	1,2,3-Trichloropropane	0.929	1.060	-14.1	96	0.00
77 T	Bromobenzene	0.925	0.997	-7.8	94	0.00
78 T	n-propylbenzene	3.767	4.098	-8.8	95	0.00
79 T	2-Chlorotoluene	2.283	2.362	-3.5	93	0.00
80 T	1,3,5-Trimethylbenzene	2.817	2.975	-5.6	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.343	-32.4#	109	0.00
82 T	4-Chlorotoluene	2.618	2.808	-7.3	95	0.00
83 T	tert-Butylbenzene	2.820	2.962	-5.0	90	0.00
84 T	1,2,4-Trimethylbenzene	2.963	3.055	-3.1	92	0.00
85 T	sec-Butylbenzene	3.457	3.553	-2.8	90	0.00
86 T	p-Isopropyltoluene	3.039	3.212	-5.7	91	0.00
87 T	1,3-Dichlorobenzene	1.667	1.786	-7.1	95	0.00
88 T	1,4-Dichlorobenzene	1.709	1.800	-5.3	95	0.00
89 T	n-Butylbenzene	2.586	2.731	-5.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX010219\
Data File : VX006871.D
Acq On : 02 Jan 2019 10:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 03 00:19:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.414	0.520	-25.6#	104	0.00
91 T	1,2-Dichlorobenzene	1.667	1.772	-6.3	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.210	0.231	-10.0	94	0.00
93 T	1,2,4-Trichlorobenzene	1.150	1.231	-7.0	94	0.00
94 T	Hexachlorobutadiene	0.542	0.518	4.4	87	0.00
95 T	Naphthalene	3.671	3.578	2.5	86	0.00
96 T	1,2,3-Trichlorobenzene	1.187	1.200	-1.1	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6