

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
 Data File : VX039568.D
 Acq On : 02 Jan 2024 13:37
 Operator : JC/MD
 Sample : 06046-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 03 04:12:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	218069	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93033	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90162	47.775	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.540%
35) Dibromofluoromethane	5.385	113	70627	46.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	272587	48.737	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.079	95	114997	49.833	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
Data File : VX039568.D
Acq On : 02 Jan 2024 13:37
Operator : JC/MD
Sample : 06046-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 03 04:12:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

