

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
 Data File : VX039569.D
 Acq On : 02 Jan 2024 14:00
 Operator : JC/MD
 Sample : 06046-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 03 04:13:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	208899	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88339	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86623	47.218	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.440%
35) Dibromofluoromethane	5.385	113	69103	47.883	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	8.647	98	261675	48.840	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	111454	50.418	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
Data File : VX039569.D
Acq On : 02 Jan 2024 14:00
Operator : JC/MD
Sample : 06046-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jan 03 04:13:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

