

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
 Data File : VX039570.D
 Acq On : 02 Jan 2024 14:23
 Operator : JC/MD
 Sample : 06046-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 03 04:13:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	89823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	192770	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79788	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79759	48.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.940%
35) Dibromofluoromethane	5.385	113	61706	46.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	241318	48.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	99486	48.769	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
Data File : VX039570.D
Acq On : 02 Jan 2024 14:23
Operator : JC/MD
Sample : 06046-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 03 04:13:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

