

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
 Data File : VX039575.D
 Acq On : 02 Jan 2024 16:31
 Operator : JC/MD
 Sample : P1002-10 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1218

Quant Time: Jan 03 04:15:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	186563	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76564	47.771	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.540%
35) Dibromofluoromethane	5.385	113	60362	46.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.660%
50) Toluene-d8	8.647	98	233161	48.728	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.460%
62) 4-Bromofluorobenzene	11.079	95	99743	50.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.040%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.782	84	701	0.525	ug/l #	90
25) 2-Butanone	4.587	43	3888	2.811	ug/l	95
52) Toluene	8.726	92	1019	0.293	ug/l	100
68) m/p-Xylenes	10.299	106	1161	0.420	ug/l	99
95) Naphthalene	13.780	128	2707	0.443	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
Data File : VX039575.D
Acq On : 02 Jan 2024 16:31
Operator : JC/MD
Sample : P1002-10 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1218

Quant Time: Jan 03 04:15:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

