

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
 Data File : VX039578.D
 Acq On : 02 Jan 2024 17:39
 Operator : JC/MD
 Sample : P1002-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1211

Quant Time: Jan 03 04:16:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	185176	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	82763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76533	48.061	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.120%
35) Dibromofluoromethane	5.385	113	60369	47.190	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	8.647	98	232363	48.925	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	99300	50.674	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.340%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	574	0.432	ug/l	87
25) 2-Butanone	4.574	43	4757	3.462	ug/l	91
68) m/p-Xylenes	10.305	106	697	0.254	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010224\
Data File : VX039578.D
Acq On : 02 Jan 2024 17:39
Operator : JC/MD
Sample : P1002-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1211

Quant Time: Jan 03 04:16:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

