

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010319\
 Data File : VX006885.D
 Acq On : 03 Jan 2019 10:59
 Operator : JC/MD
 Sample : VX0103WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0103WBL01

Quant Time: Jan 04 00:39:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	758982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1146257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1017995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	499324	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	388570	47.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.00%	
35) Dibromofluoromethane	5.51	113	352329	48.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.22%	
50) Toluene-d8	8.71	98	1330691	48.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.38%	
62) 4-Bromofluorobenzene	11.13	95	471916	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010319\
Data File : VX006885.D
Acq On : 03 Jan 2019 10:59
Operator : JC/MD
Sample : VX0103WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0103WBL01

Quant Time: Jan 04 00:39:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

