

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010319\
 Data File : VX006891.D
 Acq On : 03 Jan 2019 13:39
 Operator : JC/MD
 Sample : K1029-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B12_123118

Quant Time: Jan 04 04:35:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	664087	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1005394	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	915335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	441569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	343219	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	5.50	113	305807	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
50) Toluene-d8	8.71	98	1178637	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.13	95	421596	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
Target Compounds						
30) Chloroform	5.21	83	12744	1.128	ug/l #	Qvalue 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010319\
Data File : VX006891.D
Acq On : 03 Jan 2019 13:39
Operator : JC/MD
Sample : K1029-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B12_123118

Quant Time: Jan 04 04:35:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

