

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010319\  
 Data File : VX006892.D  
 Acq On : 03 Jan 2019 14:05  
 Operator : JC/MD  
 Sample : K1029-03  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 B1 123118

Quant Time: Jan 04 04:37:36 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X121818W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 19 15:20:58 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 676739   | 50.00 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 1030810  | 50.00 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 935481   | 50.00 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 451924   | 50.00 | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |       |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 353918   | 48.52 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 97.04% |          |
| 35) Dibromofluoromethane    | 5.51  | 113  | 311791   | 47.83 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 95.66% |          |
| 50) Toluene-d8              | 8.71  | 98   | 1198683  | 48.78 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 97.56% |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 426265   | 47.36 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 94.72% |          |
| Target Compounds            |       |      |          |       |        |          |
| 30) Chloroform              | 5.21  | 83   | 54723    | 4.754 | ug/l   | 93       |
| 64) Tetrachloroethene       | 9.33  | 164  | 6049     | 0.817 | ug/l   | 89       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010319\  
Data File : VX006892.D  
Acq On : 03 Jan 2019 14:05  
Operator : JC/MD  
Sample : K1029-03  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
B1\_123118

Quant Time: Jan 04 04:37:36 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X121818W.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 19 15:20:58 2018  
Response via : Initial Calibration

