

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010319\
 Data File : VX006902.D
 Acq On : 03 Jan 2019 18:32
 Operator : JC/MD
 Sample : K1019-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TEST-PIT-D-11-12-B

Quant Time: Jan 04 06:00:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	638640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	976216	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	917161	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	445350	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	334521	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	5.51	113	292607	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	8.71	98	1174183	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	417444	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	29834	9.434	ug/l	97
18) Methyl Acetate	2.78	43	26291	2.827	ug/l	95
20) Methylene Chloride	2.86	84	24257	3.710	ug/l	98
25) 2-Butanone	4.70	43	5231	1.176	ug/l #	89
43) Isopropyl Acetate	6.46	43	30005	2.343	ug/l	100
95) Naphthalene	13.83	128	51044	1.561	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010319\
Data File : VX006902.D
Acq On : 03 Jan 2019 18:32
Operator : JC/MD
Sample : K1019-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TEST-PIT-D-11-12-B

Quant Time: Jan 04 06:00:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

