


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010319\  
Data File : VX006883.D  
Acq On : 03 Jan 2019 09:44  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X121818W.M  
Title : SW846 8260  
Last Update : Wed Dec 19 15:20:58 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 16.8      | 16930   | PASS             |
| 75          | 95           | 30           | 60           | 48.6      | 48956   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 100800  | PASS             |
| 96          | 95           | 5            | 9            | 6.5       | 6537    | PASS             |
| 173         | 174          | 0.00         | 2            | 0.8       | 745     | PASS             |
| 174         | 95           | 50           | 100          | 88.2      | 88901   | PASS             |
| 175         | 174          | 5            | 9            | 7.3       | 6510    | PASS             |
| 176         | 174          | 95           | 101          | 100.5     | 89378   | PASS             |
| 177         | 176          | 5            | 9            | 6.2       | 5531    | PASS             |