

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010320\
 Data File : VX014404.D
 Acq On : 03 Jan 2020 16:37
 Operator : JC/SP
 Sample : K6476-06ME
 Misc : 7.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022SB455_191227-14-16ME

Quant Time: Jan 06 03:31:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	546090	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	822722	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	742567	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	373439	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	291837	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
35) Dibromofluoromethane	5.50	113	232355	46.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.92%	
50) Toluene-d8	8.71	98	978628	50.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.52%	
62) 4-Bromofluorobenzene	11.14	95	359873	50.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.12%	
Target Compounds						
31) Cyclohexane	5.56	56	25008	2.704	ug/l	94
39) Methylcyclohexane	7.45	83	73304	7.864	ug/l	94
65) Chlorobenzene	10.14	112	299452	18.966	ug/l	98
83) tert-Butylbenzene	11.76	119	34859	1.619	ug/l	77
85) sec-Butylbenzene	11.94	105	180640	7.106	ug/l	91
89) n-Butylbenzene	12.39	91	246157	12.406	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010320\
Data File : VX014404.D
Acq On : 03 Jan 2020 16:37
Operator : JC/SP
Sample : K6476-06ME
Misc : 7.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022SB455_191227-14-16ME

Quant Time: Jan 06 03:31:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

