

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010320\
 Data File : VX014405.D
 Acq On : 03 Jan 2020 17:00
 Operator : JC/SP
 Sample : K6476-07ME
 Misc : 6.26g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022SB455_191227-21-23ME

Quant Time: Jan 06 03:34:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	537158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	816383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	722438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	359069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	285547	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	5.50	113	225649	45.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.94%	
50) Toluene-d8	8.71	98	952644	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.14	95	337018	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
Target Compounds						
4) Vinyl Chloride	1.41	62	8171	1.201	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	171131	26.581	ug/l	92
44) Trichloroethene	7.20	130	270148	42.435	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010320\
Data File : VX014405.D
Acq On : 03 Jan 2020 17:00
Operator : JC/SP
Sample : K6476-07ME
Misc : 6.26µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022SB455_191227-21-23ME

Quant Time: Jan 06 03:34:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

