

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010320\
 Data File : VX014411.D
 Acq On : 03 Jan 2020 19:20
 Operator : JC/SP
 Sample : L1009-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-08-010220

Quant Time: Jan 06 03:48:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	562117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	850239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	783232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	367880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	300447	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
35) Dibromofluoromethane	5.49	113	251466	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	8.71	98	1019787	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	11.13	95	352167	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
Target Compounds						
16) Acetone	2.43	43	292571	70.737	ug/l	99
20) Methylene Chloride	2.85	84	14991	2.302	ug/l	92
43) Isopropyl Acetate	6.43	43	123151	8.638	ug/l	98
95) Naphthalene	13.83	128	64795	2.699	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010320\
Data File : VX014411.D
Acq On : 03 Jan 2020 19:20
Operator : JC/SP
Sample : L1009-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-08-010220

Quant Time: Jan 06 03:48:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

