

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010320\
 Data File : VX014414.D
 Acq On : 03 Jan 2020 20:30
 Operator : JC/SP
 Sample : PB125870ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125870ZHE#03

Quant Time: Jan 06 03:55:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	557810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	853792	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	761747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	346696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	295636	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
35) Dibromofluoromethane	5.49	113	249624	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
50) Toluene-d8	8.71	98	1015115	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	11.13	95	348163	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
Target Compounds						
16) Acetone	2.43	43	26244	6.394	ug/l	95
20) Methylene Chloride	2.85	84	15124	2.340	ug/l #	86
43) Isopropyl Acetate	6.43	43	162481	11.349	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010320\
Data File : VX014414.D
Acq On : 03 Jan 2020 20:30
Operator : JC/SP
Sample : PB125870ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125870ZHE#03

Quant Time: Jan 06 03:55:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

