

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026216.D
 Acq On : 03 Jan 2022 13:42
 Operator : JC/MD
 Sample : M5257-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P2-B

Quant Time: Jan 04 03:49:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	173995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	303118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	279286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136345	54.165	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.340%		
35) Dibromofluoromethane	5.391	113	101038	51.271	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.540%		
50) Toluene-d8	8.653	98	370177	50.501	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.000%		
62) 4-Bromofluorobenzene	11.079	95	130187	49.891	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.780%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11505	8.436	ug/l	97
20) Methylene Chloride	2.794	84	6885	2.255	ug/l	95
40) Benzene	6.044	78	11671	1.287	ug/l	97
43) Isopropyl Acetate	6.342	43	17300	2.880	ug/l	96
52) Toluene	8.720	92	7373	1.320	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026216.D
Acq On : 03 Jan 2022 13:42
Operator : JC/MD
Sample : M5257-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P2-B

Quant Time: Jan 04 03:49:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

