

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
 Data File : VX026220.D
 Acq On : 03 Jan 2022 15:14
 Operator : JC/MD
 Sample : M5257-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P3-B

Quant Time: Jan 04 03:50:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	178312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	309062	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136992	53.105	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.200%
35) Dibromofluoromethane	5.391	113	102541	51.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	8.653	98	365728	48.934	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	134515	50.558	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.120%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13744	9.833	ug/l	# 88
20) Methylene Chloride	2.794	84	7609	2.507	ug/l	87
43) Isopropyl Acetate	6.342	43	17992	2.938	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010322\
Data File : VX026220.D
Acq On : 03 Jan 2022 15:14
Operator : JC/MD
Sample : M5257-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P3-B

Quant Time: Jan 04 03:50:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

